



Reference Material Data Sheet

IAG OU-9 Separation Lake Pegmatite

International Association of Geoanalysts
13 Belvedere Close, Keyworth, Nottingham NG12 5JF, UK
e-mail iageo.ltd@ntlworld.com
Telephone +44 (0)115 9375219

Description of the reference material

IAG OU-9 is a pegmatite, from the Separation Lake Pegmatite Field, Canada and was supplied as a coarsely crushed bulk sample by Fred Breaks of the Ontario Geological Survey, Canada. The material was ground, homogenised and sealed in packets at The Open University. The test material was analysed by WDXRF at the Open University for a range of major and trace elements and the data tested for consequential degree of heterogeneity according to the Fearn test. In none of the cases for which valid data were obtained was any significant lack of homogeneity found, and therefore the sample was considered suitable for use in the GeoPT proficiency testing programme. The material was characterised as a reference material using the GeoPT23/2008 round of the International Association of Geoanalysts' GeoPT proficiency testing scheme. The Proficiency Testing Steering Committee for this round was Prof. M. Thompson (statistician), Dr P.C. Webb (results coordinator), Prof. P.J. Potts, J.S. Watson and C. Kriete.

Intended use

This reference material is designed for use by laboratories undertaking the determination of the major and trace element mass concentration fractions of silicate rocks and similar materials for the calibration of a measurement system, the assessment of a measurement procedure, assigning values to other materials, and quality control. Note that the material may be used only for a single purpose in the same measurement process. For example, it must not be used for calibration and method validation at the same time.

Minimum sample size

On the basis of the homogeneity results and an assessment of the methods used to contribute results to the GeoPT23 round, the minimum sample size recommended for use as a test portion is 0.2 g.

Period of validity

Provided the storage and handling conditions are met, this reference material is not expected to deteriorate with time. On exposure to air, the material may absorb moisture, and instructions for handling must be followed.

Storage information

Store in a sealed container in a cool dry environment.

Instructions for handling

Before any measurements are made, every portion of the test sample must be dried at 105 ± 5 °C for at least 2 hours. Avoid contamination and cross-contamination of the test material.

IAG OU-9 – Separation Lake Pegmatite

Reference values							
<i>Assigned value elemental/oxide concentrations and uncertainties from the GeoPT23 report on a dried (105 °C) basis</i>							
Oxide / element	Reference value g 100 g ⁻¹	Uncertainty g 100 g ⁻¹	n	Element	Reference value mg kg ⁻¹	Uncertainty mg kg ⁻¹	n
SiO ₂	79.50	0.17	69	Ho	0.154	0.013	34
Al ₂ O ₃	12.35	0.07	72	Lu	0.040	0.002	30
Fe ₂ O ₃ T	0.74	0.01	74	Pr	1.24	0.09	34
MnO	0.106	0.002	71	Rb	2500	45	64
CaO	0.294	0.004	70	Sb	7.67	0.42	30
Na ₂ O	4.17	0.03	71	Sc	2.77	0.13	38
K ₂ O	1.36	0.01	71	Sm	3.15	0.24	39
	mg/kg	mg/kg		Tb	0.46	0.04	38
Cs	403	12	48	Th	5.08	0.35	50
Dy	1.70	0.08	38	Tl	13.8	0.8	27
Er	0.30	0.03	34	Tm	0.049	0.003	27
Ga	56.6	1.4	56	U	4.37	0.32	52
Gd	2.53	0.16	37	Zn	28.1	0.7	64

Reference values are the GeoPT assigned values assessed from the robust statistical analysis of results submitted to the GeoPT23 round, following an assessment of both the consistency of data distribution and the agreement between methods, where possible.

Uncertainties are the robust standard deviation of the mean or median of the assigned value expanded by a coverage factor of two, and rounded up.

n is the number of laboratories reporting results for that element/oxide in the GeoPT23 round.

Fe₂O₃T is the total iron expressed as Fe₂O₃.

Assessment of reference values

The reference values were determined as a 'consensus', based on the statistical location of the participants' results to the GeoPT23 round. This location was determined as a robust mean if the distribution of results was unimodal and, outliers aside, close to symmetrical. If a slight asymmetry was apparent in a unimodal distribution, the median was chosen as an alternative. In other circumstances, notably when the number of valid results contributing to the location was less than 15 or their dispersion was unusually great, no reference value was assigned, although values may be reported as information values. These judgements were made by the IAG Proficiency Testing Steering Committee.

Metreological traceability

Traceability was not formally demonstrated for this reference material. However, traceability could be demonstrated by the laboratories participating in this round by the use of certified reference materials as calibrators or for performance assessment (although this information is not currently recorded by the GeoPT programme). Furthermore, traceability is implied by the overall agreement between datasets for individual elements/oxides submitted by laboratories that contributed to the GeoPT programme.

IAG OU-9 – Separation Lake Pegmatite								
Information values								
Oxide / element	Value g 100 g ⁻¹	Uncertainty g 100 g ⁻¹	n		Element	Value mg kg ⁻¹	Uncertainty mg kg ⁻¹	n
TiO₂	0.057	0.002	68		La	2.0	0.2	45
P₂O₅	0.030	0.002	59		Li	695	30	33
LOI	0.78	0.02	59		Nb	155	7	60
	mg kg⁻¹	mg kg⁻¹			Nd	5.1	0.5	40
Ba	8.7	0.7	52		Ta	125	7	50
Ce	7.2	0.5	45		W	6.1	0.7	28
Eu	0.05	0.01	30		Y	8.1	0.9	57
Ge	5.0	0.5	20		Yb	0.34	0.03	34
Hf	1.0	0.1	40					

Information values are 'provisional' data from the relevant GeoPT report with additional 'information' values for elements that gave a reasonably cohesive data distribution. In both cases, data distributions were judged **not** to be good enough to meet the criteria for designation as reference values. These data are provided for information purposes only and **not** for the calibration of methods or the assessment of data.

Uncertainties are the robust standard deviation of the mean or median expanded by a coverage factor of two, and rounded up.

n is the number of laboratories reporting results for that element/oxide in the GeoPT23 round.

LOI is the loss on ignition.

Reference to reference material characterisation report

Further details of the procedures used, the results, their statistical analysis and assessment, on which the property values listed in this certificate are based, can be found in the GeoPT23 report, which can be downloaded free of charge for personal use (<http://www.geoanalyst.org/index.php/proficiency-testing-proficiency-testing>).

Safety information

Silicate powders containing heavy metals can cause harm especially if inhaled or in contact with the skin. User organisations must undertake a health and safety risk assessment and ensure that the appropriate procedures are followed in the handling and use of this material. Further details are available on the relevant Material Safety Data Sheet.

Legal notice – terms and conditions

1. The IAG shall not be liable to the user of this material for loss (whether direct or indirect) of profits, business, anticipated savings or reputation or for any indirect or consequential loss or damage whatsoever even if previously advised thereof and whether arising from negligence, breach of these Terms and Conditions or howsoever occurring.
2. In any event, and notwithstanding anything contained in these Terms and Conditions, IAG's liability in contract, tort (including negligence, defamation or breach of statutory duty) or otherwise arising by reason of or in connection with these Terms and Conditions (including as a result of proficiency testing) shall be limited to the price paid for the material giving rise to such liability.
3. The IAG does not grant any warranties in relation to GeoPT products or the supply of analytical services or distribution of the proficiency test, and all other conditions, warranties, stipulations or other statements whatsoever, whether express or implied, by statute, at common law or otherwise howsoever, relating to the GeoPT products, analytical services or proficiency tests are hereby excluded. In particular, (but without limitation to the foregoing) no warranties are granted regarding the fitness for purpose, performance, use, quality or merchantability of the GeoPT products, whether express or implied, by statute, at common law or otherwise howsoever.

Revisions

This Reference Material Data Sheet is Version 1.00. Any revisions to this reference material data sheet will be made available on the IAGeo Ltd web site (www.iageo.com).

Acknowledgements

The contributions of Fred Breaks (Ontario Geological Survey, Canada) in providing the sample and Peter Webb in contributing to this data sheet are gratefully acknowledged.

Approvals

This reference material information sheet was approved on behalf of the Reference Material and Certification Committee of the International Association of Geoanalysts.

Name *Philip J. Potts*

Position *Chair of the IAG Certification
and Reference Material Committee*

Date **21st December 2015**

©2015 International Association of Geoanalysts.