



Uncertainty estimates for evaluating the significance of bias of PXRF measurements using the SdAR-H1, SdAR-M2 and SdAR-L2 reference materials

**Annex to Reference Material Data Sheets
SdAR-H1 Metalliferous sediment
SdAR-M2 Metal-rich sediment
SdAR-L2 Blended sediment**

Peter Rostron and Michael H. Ramsey
University of Sussex
January 2017

International Association of Geoanalysts
13 Belvedere Close, Keyworth, Nottingham NG12 5JF, UK
Email: iageo.ltd@ntlworld.com
Telephone: +44 (0)115 9375219

Purpose of this document

When XRF instruments such as hand-held portable XRF (PXRF) are used for analysis, the mass of the test-portion analyzed may be significantly lower than the recommended minimum mass for laboratory analysis (0.2 g) specified on the relevant data sheet. In particulate heterogeneous materials, the occurrence of small-scale heterogeneity of analytes may then lead to increased uncertainties compared with those quoted for the reference values on the data sheet. Indeed, this issue can also affect routine XRF measurements, due to the limited critical penetration depth of fluorescence X-rays from low atomic number elements.

In this document, revised estimates of the uncertainties of SdAR reference values are given for 18 oxides/elements in SdAR-H1 (Appendix 1), 17 oxides/elements in SdAR-M2 (Appendix 2) and 16 oxides/elements in SdAR-L2 (Appendix 3), for use when materials are analysed with a PXRF instrument at two different beam sizes¹. The revised uncertainty data are provided for nominal primary X-ray beam diameters of 8 mm and 3 mm, based on measurements made with a Niton PXRF model XL3t Ultra on the following presentations of the test material:

- a) Compressed powder pellets (here, 13 mm diameter x 7 mm thickness, formed from powder moistened uniformly with a few drops of 2% PVA solution at a pressure of 8 tonnes and dried)
- b) Loose powders (here, filling a 25 mm diameter sample cup to a depth of 25 mm)

The increased uncertainty in reference values requires reassessment of experimental bias evaluated using this reference material. Two worked examples of a typical evaluation are given, one where bias is found to be significant at the 95% confidence level, and one where bias is found to be not significant.

Note that other PXRF equipment may use different nominal beam sizes and there may also be differences between geometries of X-ray detection systems.

Assessing the significance of bias measurements – Basic Principles

The uncertainty on the reference value (RV) effectively expresses the limited range of measured values around the reference value that can arise without indicating significant bias in the analytical method. Both the reference value and each measurement have uncertainty and both of these uncertainties should be taken into account when evaluating whether bias is significant.

The estimation of analytical bias is normally undertaken as part of the validation of an analytical method and is usually monitored as part of routine QC. Bias can be estimated from repeated measurements (e.g. on both sides of a single pellet) made on one or more RMs that are approximately matrix matched in terms of analyte mass-fraction, bulk composition and grain size. It is not recommended to use a single measurement on one RM for the estimation of bias, because it is then impossible to provide a reliable estimate of uncertainty on the measurement result (U_m).

The principles outlined in **ERM Application Note 1²** state that bias is not considered significant if:

$$\Delta_m \leq U_\Delta$$

where Δ_m is the absolute difference between mean measured value and reference value, and U_Δ is the uncertainty on Δ_m . This is calculated as:

$$U_\Delta = \sqrt{(U_m)^2 + (U_{RV})^2}$$

U_m = uncertainty of the measurement result

U_{RV} = uncertainty of the reference value

U_m (2s) for single measurement values can be estimated using the methods described in the Eurachem/CITAC guide³, with coverage factor adjusted for the number of replicates measured, if applicable (i.e. less than 8 independent measurements). Alternatively 'n' replicate measurements can be made on the RM over a period, and the standard error of the mean value (calculated as standard deviation(s) divided by \sqrt{n}), and doubled to give an estimate of U_m , as uncertainty on the mean value (i.e. $U_m = 2s/\sqrt{n}$).

When the test portion mass is known to be above the minimum of 0.2 g then U_{RV} can be read as the 2 sigma uncertainty from the Appendices provided in the SdAR data sheets. When measurements are made using PXRF it is recommended that appropriate revised values from the Appendices presented in this document are used. These values include a measured component of uncertainty that is caused by small-scale heterogeneity of the analyte, at the specified beam size.

Worked example for deciding whether a value of bias estimated by measurement is statistically significant

Measured values for As are given for the 8 mm beam and the 3 mm beam diameters in **Table 1**. The values are measurements made, for this example, on both sides of 6 pellets made from the SdAR-H1 reference material, using the same PXRf.

Table 1 Twelve measurements of As in SdAR-H1 using 8 mm and 3 mm beam sizes

| | |
|----------------------------------|--|
| 8 mm beam (mg kg ⁻¹) | 405, 443, 438, 427, 416, 378, 426, 457, 404, 469, 416, 414 |
| 3 mm beam (mg kg ⁻¹) | 439, 372, 414, 410, 457, 369, 383, 385, 374, 456, 428, 449 |

Methods of estimating U_m are described in ERM Application Note 1². At the most basic level, uncertainty can be estimated as 2 x the standard error of a number of successive measurements of the reference material. In this case it is recommended that the reference material sample be moved to different positions between each measurement in order to incorporate heterogeneity effects into the estimate. The measured value would then be the arithmetic mean of these measurements and U_m calculated (from **Table 1**) as

$$U_m = 2 \times \frac{\text{Standard deviation of 12 measurements}}{\sqrt{12}}$$

Summary inputs to the calculation of bias significance are shown in **Table 2**. RV (Reference Value) and $U_{RV \text{ beam}}$ (Uncertainty 8 mm beam, pellet sample) have been taken from **Appendix 1**, and U_m has been calculated as above.

Table 2 Summary of inputs to bias significance calculation (As)

| | Measured mg kg ⁻¹ | U_m mg kg ⁻¹ | RV mg kg ⁻¹ | $U_{RV \text{ beam}}$ mg kg ⁻¹ |
|------|---------------------------------|------------------------------|---------------------------|--|
| 8 mm | 425 | 14 | 396 | 24 |
| 3 mm | 411 | 20 | 396 | 60 |

Bias significance calculations

8 mm beam

$$U_{\Delta} = \sqrt{(U_m^2 + U_{RV}^2)} = 28 \text{ mg kg}^{-1}$$

$$\Delta_m = | \text{measured value} - \text{RV} | = 29 \text{ mg kg}^{-1}$$

$\Delta_m > U_{\Delta}$ therefore the bias of **29 mg kg⁻¹** is considered **significant** at 95% confidence.

3 mm beam

$$U_{\Delta} = \sqrt{(U_m^2 + U_{RV}^2)} = 63 \text{ mg kg}^{-1}$$

$$\Delta_m = | \text{measured value} - RV | = 15 \text{ mg kg}^{-1}$$

$\Delta_m < U_{\Delta}$ therefore the bias of **15 mg kg⁻¹** is considered **not significant** at 95% confidence.

When multiple CRMs are used for the estimation of analytical bias as a function of mass fraction, then a model of measured versus certified value can be fitted with least-squares regression or FREML⁴. The statistical significance of the bias can be tested by inspecting the confidence limits of the slope and intercept coefficients of the model, to see whether the confidence intervals include unity or zero respectively.

References

1. Rostron P. and Ramsey, M.H. (2017), Quantifying heterogeneity of small test portion masses of geological reference materials by PXRF: implications for uncertainty of reference values. Geostandards and Geoanalytical Research, doi:10.1111/ggr.12162, onlinelibrary.wiley.com/doi/10.1111/ggr.12162/full
2. Linsinger T. (2010) Comparison of a measurement result with the certified value, European Commission - Joint Research Centre Institute for Reference Materials and Measurements (IRMM) from http://www.erm-crm.org/ERM_products/application_notes/application_note_1/Pages/index.aspx
3. Ramsey M.H. and Ellison S.L.R. (eds.) 2007 Eurachem/EUROLAB/CITAC/Nordtest/AMC Guide: Measurement uncertainty arising from sampling: a guide to methods and approaches. Eurachem (2007) Available from the Eurachem secretariat. https://eurachem.org/images/stories/Guides/pdf/UfS_2007.pdf
4. AMC (2002) Fitting a linear functional relationship to data with error on both variables. Technical Brief 10 from Analytical Methods Committee of the Royal Society of Chemistry. http://www.rsc.org/images/linear-functional-relationship-technical-brief-10_tcm18-214865.pdf

Legal notice – terms and conditions

1. The IAG shall not be liable to the user of this material for loss (whether direct or indirect) of profits, business, anticipated savings or reputation or for any indirect or consequential loss or damage whatsoever even if previously advised thereof and whether arising from negligence, breach of these Terms and Conditions or howsoever occurring.
2. In any event, and notwithstanding anything contained in these Terms and Conditions, IAG's liability in contract, tort (including negligence, defamation or breach of statutory duty) or otherwise arising by reason of or in connection with these Terms and Conditions (including as a result of proficiency testing) shall be limited to the price paid for the material giving rise to such liability.
3. The IAG does not grant any warranties in relation to GeoPT products or the supply of analytical services or distribution of the proficiency test, and all other conditions, warranties, stipulations or other statements whatsoever, whether express or implied, by statute, at common law or otherwise howsoever, relating to the GeoPT products, analytical services or proficiency tests are hereby excluded. In particular, (but without limitation to the foregoing) no warranties are granted regarding the fitness for purpose, performance, use, quality or merchantability of the GeoPT products, whether express or implied, by statute, at common law or otherwise howsoever.

This annex to the reference material information sheet is based upon published research³, and was approved on behalf of the Reference Material and Certification Committee of the International Association of Geoanalysts.

Appendix 1 - SdAR-H1 Metalliferous Sediment

| Reference values with revised uncertainty estimates | | | | | | |
|---|--|---|---|---|---|---|
| | | | 8 mm beam instrument | | 3 mm beam instrument | |
| Oxide/ Element | Reference Value g 100g ⁻¹ | Bulk uncertainty g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ |
| Al ₂ O ₃ | 11.83 | 0.07 | 0.9 | 1.3 | 0.7 | 3 |
| CaO | 1.46 | 0.01 | 0.2 | 0.01 | 0.22 | 0.07 |
| Fe ₂ O ₃ | 6.45 | 0.04 | 0.05 | 0.05 | 0.09 | 0.15 |
| K ₂ O | 4.17 | 0.03 | 0.07 | 0.07 | 0.09 | 0.15 |
| MnO | 0.515 | 0.005 | 0.012 | 0.01 | 0.03 | 0.021 |
| SiO ₂ | 65.45 | 0.18 | 3 | 1.9 | 3 | 7 |
| TiO ₂ | 0.56 | 0.004 | 0.02 | 0.03 | 0.08 | 0.06 |
| | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ |
| As* | 396 | 24 | 24 | 40 | 60 | 60 |
| Ba | 866 | 12 | 12 | 30 | 12 | 50 |
| Cu | 1160 | 20 | 40 | 21 | 50 | 50 |
| Mo | 64 | 3 | 3 | 7 | 10 | 24 |
| Nb | 21.9 | 0.9 | 2.2 | 1.2 | 11 | 4 |
| Pb | 3890 | 80 | 80 | 90 | 130 | 120 |
| Rb | 152 | 3 | 3 | 3 | 3 | 6 |
| Sr | 182 | 3 | 4 | 4 | 4 | 6 |
| V | 73.2 | 1.7 | 7 | 1.7 | 5 | 5 |
| Zn | 3680 | 60 | 109 | 90 | 170 | 180 |
| Zr* | 258 | 60 | 60 | 70 | 70 | 80 |

* Given as information values on reference material data sheet.

Revised uncertainties have been adjusted to the coverage factor for 95% confidence using the Students t-distribution for 11 degrees of freedom in the uncertainty estimates.

Appendix 2 - SdAR-M2 Metal-rich Sediment

| Reference values with revised uncertainty estimates | | | | | | |
|---|--|---|---|---|---|---|
| | | | 8 mm beam instrument | | 3 mm beam instrument | |
| Oxide/ Element | Reference Value g 100g ⁻¹ | Bulk uncertainty g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ |
| Al ₂ O ₃ | 12.47 | 0.06 | 0.7 | 0.4 | 1.3 | 0.7 |
| CaO | 0.84 | 0.01 | 0.25 | 0.01 | 0.3 | 0.06 |
| Fe ₂ O ₃ | 2.63 | 0.02 | 0.05 | 0.05 | 0.14 | 0.08 |
| K ₂ O | 5 | 0.03 | 0.09 | 0.09 | 0.11 | 0.15 |
| MnO | 0.134 | 0.002 | 0.003 | 0.008 | 0.007 | 0.016 |
| SiO ₂ | 73.45 | 0.17 | 1.7 | 1.9 | 4 | 3 |
| TiO ₂ | 0.3 | 0.003 | 0.02 | 0.019 | 0.05 | 0.04 |
| | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ |
| As* | 76 | 5 | 10 | 6 | 30 | 12 |
| Ba | 990 | 12 | 24 | 40 | 12 | 40 |
| Cu | 236 | 4 | 13 | 10 | 19 | 50 |
| Mo | 13.3 | 0.4 | 1.5 | 1.2 | 0.4 | 1.7 |
| Nb | 26.2 | 0.7 | 1.8 | 1.5 | 4 | 3 |
| Pb | 808 | 14 | 15 | 21 | 21 | 24 |
| Rb | 149 | 2 | 4 | 2.5 | 2 | 3 |
| Sr | 144 | 3 | 3 | 5 | 3 | 3 |
| Zn | 750 | 13 | 24 | 13 | 60 | 60 |
| Zr | 259 | 7 | 60 | 25 | 30 | 80 |

* Given as information values on reference material data sheet.

Revised uncertainties have been adjusted to the coverage factor for 95% confidence using the Students t-distribution for 11 degrees of freedom in the uncertainty estimates.

Appendix 3 - SdAR-L2 Blended Sediment

| Reference values with revised uncertainty estimates | | | | | | |
|---|--|---|---|---|---|---|
| | | | 8 mm beam instrument | | 3 mm beam instrument | |
| Oxide/ Element | Reference Value g 100g ⁻¹ | Bulk uncertainty g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ | Uncertainty pellet g 100g ⁻¹ | Uncertainty powder g 100g ⁻¹ |
| Al ₂ O ₃ | 11.58 | 0.05 | 0.6 | 0.5 | 1.4 | 1.4 |
| CaO | 1.06 | 0.01 | 0.17 | 0.04 | 0.19 | 0.07 |
| Fe ₂ O ₃ | 3.63 | 0.02 | 0.12 | 0.11 | 0.15 | 0.11 |
| K ₂ O | 4.1 | 0.02 | 0.05 | 0.12 | 0.12 | 0.15 |
| MnO | 0.099 | 0.001 | 0.003 | 0.003 | 0.005 | 0.009 |
| SiO ₂ | 74.48 | 0.11 | 1.8 | 2.3 | 4 | 4 |
| TiO ₂ | 0.62 | 0.003 | 0.03 | 0.04 | 0.12 | 0.07 |
| | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ | mg kg⁻¹ |
| Ba | 809 | 10 | 30 | 110 | 16 | 70 |
| Cu | 50.8 | 1.1 | 1.1 | 7 | 9 | 5 |
| Nb | 63 | 1.5 | 1.7 | 3 | 6 | 4 |
| Pb | 183 | 3 | 5 | 9 | 7 | 4 |
| Rb | 120 | 1 | 1.3 | 1.6 | 1 | 1.1 |
| Sr | 150 | 2 | 2.2 | 5 | 3 | 2.4 |
| V | 35 | 0.8 | 0.8 | 0.8 | 4 | 1.8 |
| Zn | 201 | 3 | 3 | 5 | 9 | 14 |
| Zr | 618 | 10 | 60 | 50 | 90 | 100 |

Revised uncertainties have been adjusted to the coverage factor for 95% confidence using the Students t-distribution for 11 degrees of freedom in the uncertainty estimates.